

Microcone® Implant



Quattrocone Implant

| NI | RI | | | | | RI | | | |
|---------------|----------|----------|----------|----------|---------------|----------|----------|----------|--|
| Ø 3,0 mm | Ø 3,5 mm | Ø 4,0 mm | Ø 4,5 mm | Ø 5,0 mm | Ø 3,5 mm | Ø 3,8 mm | Ø 4,3 mm | Ø 5,0 mm | |
| | | | | | | | | | |
| 6,5 mm - | - | 2-01-35 | 2-01-41 | 2-01-47 | 6,5 mm - | - | - | - | |
| 7 mm - | - | - | - | - | 7 mm - | - | 3-01-16 | 3-01-06 | |
| 8 mm - | 2-01-30 | 2-01-36 | 2-01-42 | 2-01-48 | 8 mm - | - | - | - | |
| 9 mm - | 2-01-31 | 2-01-37 | 2-01-43 | 2-01-49 | 9 mm 3-01-02 | 3-01-17 | 3-01-07 | - | |
| 11 mm 1-01-06 | 2-01-32 | 2-01-38 | 2-01-44 | 2-01-50 | 11 mm 3-01-03 | 3-01-18 | 3-01-08 | 3-01-13 | |
| 13 mm 1-01-07 | 2-01-33 | 2-01-39 | 2-01-45 | 2-01-51 | 13 mm 3-01-04 | 3-01-19 | 3-01-09 | 3-01-14 | |
| 15 mm 1-01-08 | 2-01-34 | 2-01-40 | 2-01-46 | 2-01-52 | 15 mm - | 3-01-20 | - | 3-01-15 | |


Closure screw



Gingiva former



**Auxiliary
Instruments**



Prosthetics

Implant-level Impression

Implant pick-up Open tray Impression

| | NI | RI |
|-------|---------|---------|
| short | - | 2-04-01 |
| long | 1-04-01 | 2-04-02 |

Scanbody

| | NI | RI |
|---------|----|---------|
| 1-09-03 | | 2-09-10 |

Temporary abutment

| | NI | RI |
|---------|----|---------|
| 1-17-04 | | 2-17-07 |
| 1-06-01 | | 2-06-03 |

Incl. screw

| | NI | RI |
|---------|----|---------|
| 1-05-01 | | 2-05-01 |

Analog


| | NI | RI |
|---------|----|---------|
| 1-05-01 | | 2-05-01 |

CAD/CAM Analog

| | NI | RI |
|---------|----|---------|
| 1-05-02 | | 2-05-02 |

Novaloc®

Novaloc® Processing package titanium




2010.601

Novaloc® abutment

| Gingiva height | 0° | | 15° | |
|----------------|---------|---------|---------|----|
| | RI | RI | RI | RI |
| 1,0 mm | 2-23-01 | 2-23-06 | 2-23-11 | |
| 2,0 mm | 2-23-02 | 2-23-07 | 2-23-12 | |
| 3,0 mm | 2-23-03 | 2-23-08 | 2-23-13 | |
| 4,0 mm | 2-23-04 | 2-23-09 | 2-23-14 | |
| 5,0 mm | 2-23-05 | 2-23-10 | 2-23-15 | |

Placement instrument Novaloc® abutment angled



0-13-59

Type 1 Type 2

Implant-level Cementable/Screw retained Abutments

Titanium base

| | RI | RI | RI | RI |
|-----------------|---------|---------|---------|---------|
| Chimney height | 3,5 mm | | 5,5 mm | |
| Gingiva height | 0,6 mm | 1,1 mm | 0,6 mm | 1,1 mm |
| Article No. | 2-09-11 | 2-09-13 | 2-09-12 | 2-09-14 |
| Including Screw | 2-06-03 | | | |

Titanium base Bridges/bars

| | RI |
|-----------------|---------|
| Chimney height | 3,5 mm |
| Gingiva height | 0,6 mm |
| Article No. | 2-09-15 |
| Including Screw | 2-06-03 |

Titanium base

| | NI | NI |
|-----------------|---------|---------|
| Chimney height | 3,5 mm | |
| Gingiva height | 0,7 mm | 1,2 mm |
| Article No. | 1-09-01 | 1-09-02 |
| Including Screw | 1-06-01 | |

Titanium base Bridges/bars

| | NI |
|-----------------|---------|
| Chimney height | 3,5 mm |
| Gingiva height | 0,7 mm |
| Article No. | 1-09-04 |
| Including Screw | 1-06-01 |

Recommended torque 25 Ncm

Recommended torque 15 Ncm

Abutment level Cementable/Screw retained Abutments

Multi-unit abutment

| | 0° | | | | 17° | | | | | 30° | | |
|----------------|---------|---------|---------|--------|---------|---------|---------|---------|---------|---------|----|----|
| | RI | RI | RI | | RI | RI | RI | RI | RI | RI | RI | RI |
| Gingiva height | 1,5 mm | 3,5 mm | 5,5 mm | - | 2,5 mm | 3,5 mm | 5,5 mm | 3,0 mm | 4,0 mm | 5,5 mm | | |
| Article No. | 2-31-01 | 2-31-02 | 2-31-03 | Type 1 | 2-31-04 | 2-31-05 | 2-31-06 | 2-31-07 | 2-31-08 | 2-31-09 | | |
| | | | | Type 2 | 2-31-10 | 2-31-11 | 2-31-12 | 2-31-13 | 2-31-14 | 2-31-15 | | |

Recommended torque 25 Ncm

Type 1 Type 2

Multi-unit prosthetic components


| | Titanium Cap | Titanium Base | Modelling sleeve |
|-------------|--------------|---------------|------------------|
| Article No. | 0-31-06 | 0-31-09 | 0-31-11 |
| Incl. Screw | 0-31-02 | | |

Recommended torque 15 Ncm

Multi-unit prosthetic accessories

| | Analog | Implant pick up | Multi-unit cover cap | Scanbody | Analog CAD/CAM |
|-------------|---------|-----------------|----------------------|----------|----------------|
| Article No. | 0-31-05 | 0-31-04 | 0-31-03 | 0-31-01 | 0-31-10 |

Placement instrument Multi-unit abutment straight



0-13-76

MedentiBASE abutments

| | RI | RI | RI | RI | RI |
|----------------|---------|---------|---------|---------|---------|
| Gingiva height | 0,5 mm | 1,5 mm | 2,5 mm | 3,5 mm | 4,5 mm |
| Article No. | 2-28-01 | 2-28-02 | 2-28-03 | 2-28-04 | 2-28-05 |

Recommended torque 25 Ncm

MedentiBASE prosthetic components


| | Titanium Cap | Titanium Base | Plastic Cap |
|-------------|--------------|---------------|-------------|
| Article No. | 0-28-06 | 0-28-05 | 0-28-08 |
| Incl. Screw | 0-06-03 | | |

Recommended torque 15 Ncm

MedentiBASE accessories

| | Analog | Implant pick up | Cover Cap | Scanbody | Analog CAD/CAM |
|-------------|---------|-----------------|-----------|----------|----------------|
| Article No. | 0-28-20 | 0-28-12 | 0-28-04 | 0-28-13 | 0-28-21 |

Placement instrument MedentiBASE abutments



0-13-37